

Required Information

When Requesting an offer for a Protective Curtain Model VH194-NV


0 Area of Use: Indoor X-ray protection for windows
(Not suitable for outdoor use or as protection for doors or entrances.)

1.0 Curtain Size: _____ cm x _____ cm (total height including track x complete length/track length)
(If known)
(Fill out only after **taking steps 2.1 to 4.0 into consideration.**
Steps 2.0, 5.0 to 6.2, and 7.0 still need to be completed!)

2.0 Mounted To: Ceiling or Wall

2.1 Mounting Variation: Inside window recess | Outside window recess

2.2 Further Information: _____ cm Recess height
_____ cm Recess width



2.3 Spacing **Left:** _____ cm Recess wall to window
Right: _____ cm Window to recess wall | _____ cm Room wall to window
_____ cm Window to room wall
(With enough space, the curtain will not obstruct the window in a parked state.)

3.0 Window Size: _____ cm x _____ cm (H x W) | _____ cm x _____ cm (H x W)
(Including frame)

4.0 Window Position: _____ cm Upper edge of window to upper edge of recess | _____ cm Upper edge of window to ceiling
_____ cm Lower edge of window to lower edge of recess | _____ cm Lower edge of window to floor

NEW

From June, 2022 – Optimized durability due to reinforced lead rubber panel suspension!

5.0 Curtain Pb-Value: 0.50 mm (6.80 m) | 1.00 mm (4.08 m) | 1.50 mm (2.82 m) | 2.00 mm (2.04 m)
(Maximum curtain length)

6.1 Panel Park Position: Right or Left
(When possible, the park position should be outside of the window.)

6.2 Position control Panel Right or Left

7.0 Additional Information: Please give the position and measurements of any special structural conditions, radiator, protruding window bank, etc.
(Optional)

Important Information

- Panels in working condition: For safety, there should be an overlap of 10 cm on all sides of the window.
- A recommended distance of 2-3 cm (min. 1 cm) from the bottom of the panels to the floor or window sill is needed so that the panels can properly move and turn.

MAVIG GmbH
X-Ray Protection and
Medical Suspension
Systems

Address
Stahlgruberring 5
81829 Munich
Germany

Mail Address
P.O. Box 820362
81803 Munich
Germany

Phone
+49 (0) 89 / 420 96 - 0
Fax
+49 (0) 89 / 420 96 - 200

e-Mail
info@mavig.com

Web
www.mavig.com